



Verification of Compliance

The submitted sample of the following equipment has been tested for CE marking according to the following European Directives: the EMC Directive 2014/30/EU

Applicant & Address : MICRO SENSOR CO.,LTD
No.18 Yingda Road, Baoji 721006, Shaanxi, P.R.China

Manufacturer & Address : MICRO SENSOR CO.,LTD
No.18 Yingda Road, Baoji 721006, Shaanxi, P.R.China

Product : Pressure Transmitter

Model/Type reference : MPM483E

Trade mark : microsensor

Report No. : EED3908012057 & EED3908012058

Test Standards	EN 61000-6-4:2007+A1:2011	
	EN 61000-3-2:2019	
	EN 61000-3-3:2013+A1:2019	
	EN IEC 61000-6-2: 2019	
Ratings	Parameter	Documentary Evidence
	Input: DC 11V~28V	EED390812057
	Output: DC 4mA-20mA	
	Input: DC 15V~28V	EED390812058
	Output: DC 0V-10VDC	

This Verification is for the exclusive use of CTI's Client and is provided pursuant to the agreement between CTI and its Client. The observations and test results referenced from this Verification are relevant only to the sample tested. This Verification by itself does not imply that the material, product, or service is or has ever been under a CTI certification program.

Note 1: This verification is part of the full test report(s) and should be read in conjunction with it.



Jeff fang

Date: Aug 29,2022

Check No.: :8905080822

CENTRE TESTING INTERNATIONAL GROUP CO., LTD.

www.cti-cert.com E-mail:info@cti-cert.com

Hotline
400-6788-333